

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100220\
 Data File : VN063821.D
 Acq On : 2 Oct 2020 15:09
 Operator : JC/MD
 Sample : PB132030TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132030TB

Quant Time: Oct 05 05:29:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	217881	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	403837	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	432785	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	179814	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	188184	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
35) Dibromofluoromethane	7.55	113	135555	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
50) Toluene-d8	10.06	98	525262	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
62) 4-Bromofluorobenzene	12.38	95	218510	54.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.44%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.73	59	2241	5.639	ug/l	97
16) Acetone	3.78	43	20025	18.072	ug/l	93
20) Methylene Chloride	4.51	84	19455	7.749	ug/l	96
25) 2-Butanone	6.79	43	15644	9.958	ug/l	98
43) Isopropyl Acetate	8.13	43	69477	10.363	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100220\
Data File : VN063821.D
Acq On : 2 Oct 2020 15:09
Operator : JC/MD
Sample : PB132030TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132030TB

Quant Time: Oct 05 05:29:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

