

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100220\
 Data File : VN063823.D
 Acq On : 2 Oct 2020 15:56
 Operator : JC/MD
 Sample : PB132030ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132030ZHE#02

Quant Time: Oct 05 06:00:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	217077	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	408560	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	428178	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	172465	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	189242	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
35) Dibromofluoromethane	7.55	113	132392	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
50) Toluene-d8	10.06	98	526677	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
62) 4-Bromofluorobenzene	12.38	95	211840	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
Target Compounds						
16) Acetone	3.78	43	24175	21.898	ug/l	95
18) Methyl Acetate	4.28	43	4126	1.439	ug/l #	88
20) Methylene Chloride	4.50	84	6769	2.770	ug/l	91
43) Isopropyl Acetate	8.13	43	80148	11.816	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100220\
Data File : VN063823.D
Acq On : 2 Oct 2020 15:56
Operator : JC/MD
Sample : PB132030ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132030ZHE#02

Quant Time: Oct 05 06:00:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

