

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100220\
 Data File : VN063826.D
 Acq On : 2 Oct 2020 17:08
 Operator : JC/MD
 Sample : PB132030ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132030ZHE#05

Quant Time: Oct 05 06:15:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	213938	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	399318	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	415407	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	169341	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	185052	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
35) Dibromofluoromethane	7.55	113	132190	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
50) Toluene-d8	10.06	98	515172	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
62) 4-Bromofluorobenzene	12.38	95	204836	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
Target Compounds						
16) Acetone	3.78	43	18691	17.179	ug/l	96
20) Methylene Chloride	4.50	84	20309	8.232	ug/l	94
25) 2-Butanone	6.79	43	15842	10.270	ug/l #	89
43) Isopropyl Acetate	8.13	43	65261	9.844	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100220\
Data File : VN063826.D
Acq On : 2 Oct 2020 17:08
Operator : JC/MD
Sample : PB132030ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132030ZHE#05

Quant Time: Oct 05 06:15:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

