

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100220\
 Data File : VN063830.D
 Acq On : 2 Oct 2020 18:44
 Operator : JC/MD
 Sample : PB132030ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132030ZHE#09

Quant Time: Oct 05 06:23:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	214602	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	408892	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	426015	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	174314	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	192392	54.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.54%	
35) Dibromofluoromethane	7.55	113	135753	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
50) Toluene-d8	10.06	98	526544	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
62) 4-Bromofluorobenzene	12.38	95	207213	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
Target Compounds						
16) Acetone	3.78	43	11831	10.840	ug/l	96
18) Methyl Acetate	4.29	43	3615	1.276	ug/l #	68
20) Methylene Chloride	4.50	84	10676	4.360	ug/l	95
43) Isopropyl Acetate	8.13	43	76583	11.281	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100220\
Data File : VN063830.D
Acq On : 2 Oct 2020 18:44
Operator : JC/MD
Sample : PB132030ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132030ZHE#09

Quant Time: Oct 05 06:23:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

