

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100220\
 Data File : VN063833.D
 Acq On : 2 Oct 2020 19:56
 Operator : JC/MD
 Sample : PB132030ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132030ZHE#12

Quant Time: Oct 05 06:30:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	209777	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	409567	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	413109	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	170029	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	188813	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
35) Dibromofluoromethane	7.55	113	133572	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
50) Toluene-d8	10.06	98	516833	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
62) 4-Bromofluorobenzene	12.38	95	203552	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
Target Compounds						
16) Acetone	3.78	43	12970	12.157	ug/l #	82
20) Methylene Chloride	4.50	84	11089	4.627	ug/l #	88
43) Isopropyl Acetate	8.13	43	75489	11.102	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100220\
Data File : VN063833.D
Acq On : 2 Oct 2020 19:56
Operator : JC/MD
Sample : PB132030ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132030ZHE#12

Quant Time: Oct 05 06:30:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

