

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100220\
 Data File : VN063846.D
 Acq On : 3 Oct 2020 2:21
 Operator : JC/MD
 Sample : L4198-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RO-15

Quant Time: Oct 05 08:18:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	210639	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	397939	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	404606	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	165357	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	189590	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
35) Dibromofluoromethane	7.55	113	132226	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
50) Toluene-d8	10.06	98	518981	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	12.38	95	198087	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
Target Compounds						
16) Acetone	3.78	43	13847	12.926	ug/l	96
20) Methylene Chloride	4.51	84	9263	3.866	ug/l	94
43) Isopropyl Acetate	8.13	43	77181	11.682	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100220\
Data File : VN063846.D
Acq On : 3 Oct 2020 2:21
Operator : JC/MD
Sample : L4198-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RO-15

Quant Time: Oct 05 08:18:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

