

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100220\
 Data File : VN063847.D
 Acq On : 3 Oct 2020 2:44
 Operator : JC/MD
 Sample : L4198-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RO-21

Quant Time: Oct 05 08:19:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	210994	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	409546	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	412875	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	161923	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	188517	54.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.16%	
35) Dibromofluoromethane	7.55	113	133036	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
50) Toluene-d8	10.06	98	513488	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
62) 4-Bromofluorobenzene	12.38	95	195733	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
Target Compounds						
16) Acetone	3.78	43	12733	11.866	ug/l	94
20) Methylene Chloride	4.51	84	9944	4.136	ug/l	93
43) Isopropyl Acetate	8.13	43	73399	10.795	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100220\
Data File : VN063847.D
Acq On : 3 Oct 2020 2:44
Operator : JC/MD
Sample : L4198-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RO-21

Quant Time: Oct 05 08:19:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

