

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100220\
 Data File : VN063848.D
 Acq On : 3 Oct 2020 3:08
 Operator : JC/MD
 Sample : L4198-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RO-25

Manual Integrations
 APPROVED

MMDadoda
 10/5/2020 5:35:42 PM

Quant Time: Oct 05 08:22:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	203895	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	387708	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	403394	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	153083	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	182507	54.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.36%	
35) Dibromofluoromethane	7.55	113	130327	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
50) Toluene-d8	10.06	98	505413	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
62) 4-Bromofluorobenzene	12.38	95	188488	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
Target Compounds						
16) Acetone	3.78	43	11283	10.881	ug/l	98
18) Methyl Acetate	4.28	43	3944m	1.465	ug/l	
20) Methylene Chloride	4.51	84	9575	4.122	ug/l	88
43) Isopropyl Acetate	8.13	43	71726	11.143	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100220\
Data File : VN063848.D
Acq On : 3 Oct 2020 3:08
Operator : JC/MD
Sample : L4198-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RO-25

Manual Integrations
APPROVED

MMDadoda
10/5/2020 5:35:42 PM

Quant Time: Oct 05 08:22:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

