

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100220\
 Data File : VN063856.D
 Acq On : 3 Oct 2020 6:20
 Operator : JC/MD
 Sample : L4199-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 11940

Quant Time: Oct 05 08:37:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 01:46:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	209854	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	399720	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	416022	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	165428	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	185642	53.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.10%	
35) Dibromofluoromethane	7.55	113	133597	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
50) Toluene-d8	10.06	98	513479	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
62) 4-Bromofluorobenzene	12.38	95	197470	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
Target Compounds						
16) Acetone	3.78	43	14985	14.041	ug/l	89
18) Methyl Acetate	4.29	43	4043	1.459	ug/l #	84
20) Methylene Chloride	4.50	84	10525	4.395	ug/l	92
43) Isopropyl Acetate	8.13	43	65033	9.800	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100220\
Data File : VN063856.D
Acq On : 3 Oct 2020 6:20
Operator : JC/MD
Sample : L4199-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
11940

Quant Time: Oct 05 08:37:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N093020W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 01:46:38 2020
Response via : Initial Calibration

