

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
 Data File : VN084258.D
 Acq On : 02 Oct 2024 11:49
 Operator : JC\MD
 Sample : VN1002MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1002MBL01

Quant Time: Oct 03 01:56:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156268	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	271625	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	232144	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118560	51.170	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.340%	
35) Dibromofluoromethane	8.171	113	90862	50.740	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.480%	
50) Toluene-d8	10.565	98	328411	49.848	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.700%	
62) 4-Bromofluorobenzene	12.847	95	101264	42.190	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	84.380%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
Data File : VN084258.D
Acq On : 02 Oct 2024 11:49
Operator : JC\MD
Sample : VN1002MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1002MBL01

Quant Time: Oct 03 01:56:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

