

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
 Data File : VN084262.D
 Acq On : 02 Oct 2024 14:09
 Operator : JC\MD
 Sample : P4241-03DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TT189D1-GW-20240927DL

Quant Time: Oct 03 01:59:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	267268	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	220885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	83077	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	113198	48.839	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.680%
35) Dibromofluoromethane	8.165	113	88235	50.077	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.160%
50) Toluene-d8	10.565	98	316816	48.872	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.740%
62) 4-Bromofluorobenzene	12.847	95	103411	43.787	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.580%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.365	101	4551	2.494	ug/l #	84
44) Trichloroethene	9.353	130	71914	38.613	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
Data File : VN084262.D
Acq On : 02 Oct 2024 14:09
Operator : JC\MD
Sample : P4241-03DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TT189D1-GW-20240927DL

Quant Time: Oct 03 01:59:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

