

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
 Data File : VN084267.D
 Acq On : 02 Oct 2024 16:09
 Operator : JC\MD
 Sample : P4226-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW11-MW01I-20240924

Quant Time: Oct 03 02:01:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164287	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	281618	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	240516	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	95191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	120504	49.471	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.940%
35) Dibromofluoromethane	8.165	113	94701	51.008	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.020%
50) Toluene-d8	10.565	98	335829	49.165	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.340%
62) 4-Bromofluorobenzene	12.847	95	112931	45.382	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
Data File : VN084267.D
Acq On : 02 Oct 2024 16:09
Operator : JC\MD
Sample : P4226-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW11-MW01I-20240924

Quant Time: Oct 03 02:01:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

