

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
 Data File : VN084272.D
 Acq On : 02 Oct 2024 18:09
 Operator : JC\MD
 Sample : P4226-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW9-MW-01S-20240925

Quant Time: Oct 03 02:03:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157582	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	273445	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	233902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	93275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116483	49.855	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.700%		
35) Dibromofluoromethane	8.165	113	91248	50.617	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.240%		
50) Toluene-d8	10.565	98	332282	50.100	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.200%		
62) 4-Bromofluorobenzene	12.847	95	108440	44.880	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	89.760%		
Target Compounds						
44) Trichloroethene	9.359	130	2634	1.382	ug/l	Qvalue 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
Data File : VN084272.D
Acq On : 02 Oct 2024 18:09
Operator : JC\MD
Sample : P4226-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW9-MW-01S-20240925

Quant Time: Oct 03 02:03:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

