

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
 Data File : VN084273.D
 Acq On : 02 Oct 2024 18:33
 Operator : JC\MD
 Sample : P4226-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW9-MW-01D1-20240925

Quant Time: Oct 03 02:03:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	271415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	236062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	95512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117735	50.300	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.600%		
35) Dibromofluoromethane	8.165	113	91954	51.390	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.780%		
50) Toluene-d8	10.565	98	331684	50.384	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.760%		
62) 4-Bromofluorobenzene	12.847	95	111470	46.479	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 92.960%		
Target Compounds						
44) Trichloroethene	9.353	130	1441	0.762	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
Data File : VN084273.D
Acq On : 02 Oct 2024 18:33
Operator : JC\MD
Sample : P4226-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW9-MW-01D1-20240925

Quant Time: Oct 03 02:03:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

