

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
 Data File : VN084274.D
 Acq On : 02 Oct 2024 18:57
 Operator : JC\MD
 Sample : P4226-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW9-MW-01D2-20240925

Quant Time: Oct 03 02:03:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	159736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	283406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244524	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	96610	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121631	51.356	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.720%	
35) Dibromofluoromethane	8.165	113	93758	50.181	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.360%	
50) Toluene-d8	10.565	98	343085	49.910	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.820%	
62) 4-Bromofluorobenzene	12.847	95	115108	45.965	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	91.920%	
Target Compounds						
44) Trichloroethene	9.359	130	859	0.435	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
Data File : VN084274.D
Acq On : 02 Oct 2024 18:57
Operator : JC\MD
Sample : P4226-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW9-MW-01D2-20240925

Quant Time: Oct 03 02:03:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

