

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
 Data File : VN084277.D
 Acq On : 02 Oct 2024 20:09
 Operator : JC\MD
 Sample : P4226-22
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE117D2-20240926

Quant Time: Oct 03 02:05:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	159191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	275753	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	243919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	119663	50.698	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 101.400%		
35) Dibromofluoromethane	8.165	113	94165	51.798	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.600%		
50) Toluene-d8	10.565	98	340855	50.962	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.920%		
62) 4-Bromofluorobenzene	12.847	95	112343	46.106	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 92.220%		
Target Compounds						
44) Trichloroethene	9.347	130	1103	0.574	ug/l	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100224\
Data File : VN084277.D
Acq On : 02 Oct 2024 20:09
Operator : JC\MD
Sample : P4226-22
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE117D2-20240926

Quant Time: Oct 03 02:05:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

