

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100319\
 Data File : VN058509.D
 Acq On : 3 Oct 2019 14:59
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 04 04:40:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	332838	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	552774	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	471746	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	175263	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	250871	51.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.16%	
35) Dibromofluoromethane	7.57	113	174167	51.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.82%	
50) Toluene-d8	10.08	98	691485	52.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.20%	
62) 4-Bromofluorobenzene	12.41	95	208754	43.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.00%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100319\
Data File : VN058509.D
Acq On : 3 Oct 2019 14:59
Operator : JC/SP
Sample : IBLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 04 04:40:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

