


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN100319\
Data File : VN058501.D
Acq On : 3 Oct 2019 9:30
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Title : SW846 8260
Last Update : Mon Sep 30 08:07:38 2019



AutoFind: Scans 3306, 3307, 3308; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	11587	PASS
75	95	30	60	55.5	28570	PASS
95	95	100	100	100.0	51490	PASS
96	95	5	9	6.9	3559	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.7	38976	PASS
175	174	5	9	8.1	3142	PASS
176	174	95	101	97.3	37930	PASS
177	176	5	9	7.7	2928	PASS