

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100322\
 Data File : VN074712.D
 Acq On : 03 Oct 2022 16:01
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 04 06:10:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 05:54:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.939	168	204754	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.827	114	362390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.621	117	371066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	173332	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.292	65	272888	53.038	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.080%
35) Dibromofluoromethane	7.863	113	199798	51.286	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.580%
50) Toluene-d8	10.310	98	829097	51.777	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.560%
62) 4-Bromofluorobenzene	12.604	95	200335	37.726	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	75.460%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 04 06:10:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 05:54:56 2022
Response via : Initial Calibration

