

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100322\
 Data File : VN074717.D
 Acq On : 03 Oct 2022 19:51
 Operator : JC\MD
 Sample : N4933-02DL 4X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-DUP-20220927DL

Quant Time: Oct 04 06:14:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 05:54:56 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.933	168	203792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.827	114	363900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.615	117	369232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.545	152	177088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.286	65	271131	52.946	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.900%	
35) Dibromofluoromethane	7.863	113	198239	50.674	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.340%	
50) Toluene-d8	10.310	98	804717	50.046	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.100%	
62) 4-Bromofluorobenzene	12.609	95	219035	40.712	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 81.420%#	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.045	101	25284	6.863	ug/l #	60
12) 1,1-Dichloroethene	4.022	96	17772	4.847	ug/l #	94
44) Trichloroethene	9.074	130	384468	89.647	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100322\
Data File : VN074717.D
Acq On : 03 Oct 2022 19:51
Operator : JC\MD
Sample : N4933-02DL 4X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-DUP-20220927DL

Quant Time: Oct 04 06:14:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 05:54:56 2022
Response via : Initial Calibration

