

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100322\
 Data File : VN074718.D
 Acq On : 03 Oct 2022 20:15
 Operator : JC\MD
 Sample : N4933-04DL 4X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-678-680DL

Quant Time: Oct 04 06:14:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 05:54:56 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.940	168	189627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	336404	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.622	117	339326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	160443	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.287	65	261916	54.967	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.940%	
35) Dibromofluoromethane	7.863	113	184316	50.966	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.940%	
50) Toluene-d8	10.310	98	738619	49.689	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.380%	
62) 4-Bromofluorobenzene	12.604	95	210847	42.224	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 84.440%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.052	101	22441	6.547	ug/l #	68
12) 1,1-Dichloroethene	4.016	96	16005	4.691	ug/l	97
44) Trichloroethene	9.075	130	342806	86.466	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100322\
Data File : VN074718.D
Acq On : 03 Oct 2022 20:15
Operator : JC\MD
Sample : N4933-04DL 4X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-678-680DL

Quant Time: Oct 04 06:14:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 05:54:56 2022
Response via : Initial Calibration

