

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100322\
 Data File : VN074719.D
 Acq On : 03 Oct 2022 20:39
 Operator : JC\MD
 Sample : N4933-03DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-658-660DL

Quant Time: Oct 04 06:15:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100322W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 05:54:56 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.934	168	185712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.828	114	321900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.622	117	329973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.551	152	156582	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.286	65	249299	53.422	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.840%	
35) Dibromofluoromethane	7.869	113	182766	52.815	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.620%	
50) Toluene-d8	10.310	98	723887	50.892	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.780%	
62) 4-Bromofluorobenzene	12.604	95	205468	42.926	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 85.860%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.046	101	10100	3.009	ug/l	87
44) Trichloroethene	9.075	130	272120	71.730	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100322\
Data File : VN074719.D
Acq On : 03 Oct 2022 20:39
Operator : JC\MD
Sample : N4933-03DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-658-660DL

Quant Time: Oct 04 06:15:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100322W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 05:54:56 2022
Response via : Initial Calibration

