

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
 Data File : VN084284.D
 Acq On : 03 Oct 2024 09:18
 Operator : JC\MD
 Sample : VN1003MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1003MBL01

Quant Time: Oct 04 01:28:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	158087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	274878	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	235277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	85994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	118649	50.620	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 101.240%		
35) Dibromofluoromethane	8.171	113	94393	52.088	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.180%		
50) Toluene-d8	10.565	98	332741	49.907	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.820%		
62) 4-Bromofluorobenzene	12.847	95	99544	40.983	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 81.960%		

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
Data File : VN084284.D
Acq On : 03 Oct 2024 09:18
Operator : JC\MD
Sample : VN1003MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1003MBL01

Quant Time: Oct 04 01:28:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

