

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
 Data File : VN084285.D
 Acq On : 03 Oct 2024 09:42
 Operator : JC\MD
 Sample : VN1003WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1003WBL01

Quant Time: Oct 04 01:29:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	163611	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	286487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	249556	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	98264	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125911	51.904	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.800%
35) Dibromofluoromethane	8.165	113	98018	51.897	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.800%
50) Toluene-d8	10.565	98	350297	50.412	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.820%
62) 4-Bromofluorobenzene	12.847	95	117713	46.499	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
Data File : VN084285.D
Acq On : 03 Oct 2024 09:42
Operator : JC\MD
Sample : VN1003WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1003WBL01

Quant Time: Oct 04 01:29:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

