

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
 Data File : VN084290.D
 Acq On : 03 Oct 2024 11:52
 Operator : JC\MD
 Sample : P4226-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT205S1-20240925

Quant Time: Oct 04 01:32:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	292991	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	251495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	101020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	123671	49.158	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.320%
35) Dibromofluoromethane	8.165	113	97705	50.583	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.160%
50) Toluene-d8	10.565	98	358608	50.462	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.920%
62) 4-Bromofluorobenzene	12.847	95	119499	46.157	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.320%
Target Compounds						
30) Chloroform	7.977	83	1320	0.331	ug/l #	68
44) Trichloroethene	9.359	130	12435	6.091	ug/l	90
64) Tetrachloroethene	11.106	164	507	0.289	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
Data File : VN084290.D
Acq On : 03 Oct 2024 11:52
Operator : JC\MD
Sample : P4226-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT205S1-20240925

Quant Time: Oct 04 01:32:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

