

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
 Data File : VN084297.D
 Acq On : 03 Oct 2024 14:39
 Operator : JC\MD
 Sample : P4226-23
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW4-20240927

Quant Time: Oct 04 01:35:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	156008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	270884	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	234558	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92894	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	117432	50.768	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	101.540%	
35) Dibromofluoromethane	8.165	113	90548	50.703	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.400%	
50) Toluene-d8	10.565	98	336435	51.206	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.420%	
62) 4-Bromofluorobenzene	12.847	95	111110	46.419	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	92.840%	
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.377	101	13850	7.605	ug/l	Qvalue 100
12) 1,1-Dichloroethene	4.347	96	1043	0.594	ug/l #	56
27) cis-1,2-Dichloroethene	7.482	96	2275	1.023	ug/l	68
44) Trichloroethene	9.353	130	774899	410.515	ug/l	97
64) Tetrachloroethene	11.106	164	8835	5.408	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
Data File : VN084297.D
Acq On : 03 Oct 2024 14:39
Operator : JC\MD
Sample : P4226-23
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW4-20240927

Quant Time: Oct 04 01:35:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

