

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
 Data File : VN084302.D
 Acq On : 03 Oct 2024 16:39
 Operator : JC\MD
 Sample : P4226-23DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW4-20240927DL

Quant Time: Oct 04 01:38:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	271690	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	227741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	86437	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	114953	49.280	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.560%		
35) Dibromofluoromethane	8.165	113	91223	50.930	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.860%		
50) Toluene-d8	10.565	98	330038	50.083	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.160%		
62) 4-Bromofluorobenzene	12.847	95	108294	45.109	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	90.220%		
Target Compounds						
44) Trichloroethene	9.353	130	73417	38.778	ug/l	99
64) Tetrachloroethene	11.106	164	718	0.453	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
Data File : VN084302.D
Acq On : 03 Oct 2024 16:39
Operator : JC\MD
Sample : P4226-23DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW4-20240927DL

Quant Time: Oct 04 01:38:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

