

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
 Data File : VN084304.D
 Acq On : 03 Oct 2024 17:27
 Operator : JC\MD
 Sample : P4258-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Chamberlain Ave

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/04/2024
 Supervised By : Mahesh Dadoda 10/04/2024

Quant Time: Oct 04 01:39:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164187	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302561	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265786	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	107062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	121347	49.847	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.700%		
35) Dibromofluoromethane	8.165	113	96852	48.555	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.120%		
50) Toluene-d8	10.565	98	373293	50.867	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.740%		
62) 4-Bromofluorobenzene	12.847	95	126134	47.179	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	94.360%		
Target Compounds						
16) Acetone	4.430	43	15883	15.188	ug/l	92
20) Methylene Chloride	5.283	84	4056	1.908	ug/l #	89
43) Isopropyl Acetate	8.688	43	181785m	31.223	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100324\
Data File : VN084304.D
Acq On : 03 Oct 2024 17:27
Operator : JC\MD
Sample : P4258-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Chamberlain Ave

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/04/2024
Supervised By :Mahesh Dadoda 10/04/2024

Quant Time: Oct 04 01:39:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

