

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
 Data File : VN088014.D
 Acq On : 03 Oct 2025 10:50
 Operator : JC\MD
 Sample : VN1003MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1003MBL01

Quant Time: Oct 04 02:52:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	156171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	355926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	342908	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	154810	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	160635	61.119	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.240%
35) Dibromofluoromethane	8.147	113	137978	53.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.780%
50) Toluene-d8	10.547	98	461978	50.836	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.680%
62) 4-Bromofluorobenzene	12.829	95	150151	46.224	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
Data File : VN088014.D
Acq On : 03 Oct 2025 10:50
Operator : JC\MD
Sample : VN1003MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1003MBL01

Quant Time: Oct 04 02:52:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

