

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
 Data File : VN088019.D
 Acq On : 03 Oct 2025 13:21
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 04 02:55:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	193758	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	428468	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	445314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	224602	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	166939	51.196	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.400%
35) Dibromofluoromethane	8.147	113	144763	46.534	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	10.547	98	503670	46.040	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.080%
62) 4-Bromofluorobenzene	12.829	95	195764	50.062	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
Data File : VN088019.D
Acq On : 03 Oct 2025 13:21
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 04 02:55:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

