

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
 Data File : VN088020.D
 Acq On : 03 Oct 2025 13:57
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 04 02:55:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	169969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	375483	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	375101	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	180574	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	168389	58.868	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.740%
35) Dibromofluoromethane	8.147	113	143073	52.480	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.960%
50) Toluene-d8	10.547	98	491875	51.306	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.620%
62) 4-Bromofluorobenzene	12.829	95	169717	49.526	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
Data File : VN088020.D
Acq On : 03 Oct 2025 13:57
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 04 02:55:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

