

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
 Data File : VN088013.D
 Acq On : 03 Oct 2025 10:17
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 04 02:51:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
 Supervised By :Semsettin Yesilyurt 10/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	211900	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	407793	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	399348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	212700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	189558	53.155	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.320%			
35) Dibromofluoromethane	8.147	113	147566	49.840	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.680%			
50) Toluene-d8	10.547	98	515889	49.548	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.100%			
62) 4-Bromofluorobenzene	12.829	95	180495	48.498	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 97.000%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	119124	48.543	ug/l	99
3) Chloromethane	2.383	50	124349	49.489	ug/l	100
4) Vinyl Chloride	2.536	62	129985	50.707	ug/l	97
5) Bromomethane	2.971	94	81460	51.258	ug/l	92
6) Chloroethane	3.142	64	94576	56.742	ug/l	94
7) Trichlorofluoromethane	3.512	101	226552	54.280	ug/l	97
8) Diethyl Ether	3.965	74	85426	54.177	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.371	101	118105	53.539	ug/l	99
10) Methyl Iodide	4.577	142	110011	55.410	ug/l	95
11) Tert butyl alcohol	5.518	59	120578	234.750	ug/l	100
12) 1,1-Dichloroethene	4.341	96	118331	53.163	ug/l	92
13) Acrolein	4.177	56	101381	153.827	ug/l	99
14) Allyl chloride	5.018	41	169750	46.845	ug/l	94
15) Acrylonitrile	5.706	53	449175	278.787	ug/l	99
16) Acetone	4.418	43	356272	242.815	ug/l	97
17) Carbon Disulfide	4.712	76	325539	46.810	ug/l	100
18) Methyl Acetate	5.018	43	241763	65.840	ug/l	99
19) Methyl tert-butyl Ether	5.788	73	494136	56.830	ug/l	100
20) Methylene Chloride	5.265	84	163946	55.203	ug/l	97
21) trans-1,2-Dichloroethene	5.777	96	145759	53.418	ug/l	87
22) Diisopropyl ether	6.653	45	505565	59.764	ug/l	96
23) Vinyl Acetate	6.588	43	2159268	303.155	ug/l	99
24) 1,1-Dichloroethane	6.553	63	278930	56.439	ug/l	98
25) 2-Butanone	7.465	43	599104	267.143	ug/l	99
26) 2,2-Dichloropropane	7.471	77	140646	32.031	ug/l	95
27) cis-1,2-Dichloroethene	7.471	96	180488	55.591	ug/l	94
28) Bromochloromethane	7.800	49	129042	48.057	ug/l	97
29) Tetrahydrofuran	7.824	42	404402	281.618	ug/l	99
30) Chloroform	7.947	83	309633	57.299	ug/l	96
31) Cyclohexane	8.241	56	199447	49.875	ug/l	96
32) 1,1,1-Trichloroethane	8.147	97	252160	55.224	ug/l	97
36) 1,1-Dichloropropene	8.353	75	184552	51.838	ug/l	98
37) Ethyl Acetate	7.547	43	252668	54.045	ug/l	100
38) Carbon Tetrachloride	8.347	117	214305	52.458	ug/l	96
39) Methylcyclohexane	9.582	83	186475	47.480	ug/l	93
40) Benzene	8.588	78	631046	53.533	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
 Data File : VN088013.D
 Acq On : 03 Oct 2025 10:17
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
 Supervised By :Semsettin Yesilyurt 10/06/2025

Quant Time: Oct 04 02:51:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 05:05:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.759	41	127328	52.700	ug/l	96
42) 1,2-Dichloroethane	8.653	62	228809	52.013	ug/l	99
43) Isopropyl Acetate	8.671	43	399665	54.310	ug/l	99
44) Trichloroethene	9.329	130	142273	50.528	ug/l	94
45) 1,2-Dichloropropane	9.600	63	160540	54.312	ug/l	98
46) Dibromomethane	9.688	93	128880	55.256	ug/l	93
47) Bromodichloromethane	9.871	83	252397	55.172	ug/l	93
48) Methyl methacrylate	9.665	41	187451	55.005	ug/l	97
49) 1,4-Dioxane	9.682	88	45542	904.603	ug/l #	84
51) 4-Methyl-2-Pentanone	10.423	43	1297788	282.622	ug/l	100
52) Toluene	10.606	92	389751	52.688	ug/l	96
53) t-1,3-Dichloropropene	10.818	75	231861	49.522	ug/l	99
54) cis-1,3-Dichloropropene	10.294	75	233459	48.009	ug/l	98
55) 1,1,2-Trichloroethane	10.994	97	165143	54.906	ug/l	97
56) Ethyl methacrylate	10.853	69	250497	57.638	ug/l	97
57) 1,3-Dichloropropane	11.141	76	270690	54.717	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.141	63	726559	355.558	ug/l	100
59) 2-Hexanone	11.176	43	878869	289.453	ug/l	99
60) Dibromochloromethane	11.341	129	194024	54.865	ug/l	99
61) 1,2-Dibromoethane	11.447	107	173418	54.463	ug/l	99
64) Tetrachloroethene	11.082	164	126880	52.158	ug/l	98
65) Chlorobenzene	11.870	112	441827	51.989	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.941	131	162876	53.146	ug/l	96
67) Ethyl Benzene	11.941	91	744976	52.535	ug/l	97
68) m/p-Xylenes	12.047	106	578990	106.758	ug/l	98
69) o-Xylene	12.376	106	279227	53.589	ug/l	99
70) Styrene	12.388	104	490624	54.825	ug/l	99
71) Bromoform	12.559	173	132671	53.359	ug/l #	100
73) Isopropylbenzene	12.676	105	710671	53.910	ug/l	100
74) N-amyl acetate	12.506	43	225597m	51.938	ug/l	
75) 1,1,2,2-Tetrachloroethane	12.917	83	254009	53.479	ug/l	99
76) 1,2,3-Trichloropropane	12.970	75	197976m	47.218	ug/l	
77) Bromobenzene	12.959	156	183930	52.440	ug/l	98
78) n-propylbenzene	13.011	91	871697	52.889	ug/l	100
79) 2-Chlorotoluene	13.100	91	519382	51.754	ug/l	99
80) 1,3,5-Trimethylbenzene	13.147	105	605591	53.110	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.717	75	65552	43.414	ug/l	98
82) 4-Chlorotoluene	13.200	91	542423	53.641	ug/l	100
83) tert-Butylbenzene	13.417	119	521941	54.079	ug/l	97
84) 1,2,4-Trimethylbenzene	13.459	105	629394	53.814	ug/l	99
85) sec-Butylbenzene	13.594	105	749419	52.962	ug/l	98
86) p-Isopropyltoluene	13.706	119	629181	52.499	ug/l	99
87) 1,3-Dichlorobenzene	13.711	146	357765	52.270	ug/l	99
88) 1,4-Dichlorobenzene	13.788	146	366323	52.078	ug/l	97
89) n-Butylbenzene	14.035	91	565873	49.962	ug/l	100
90) Hexachloroethane	14.311	117	129377	51.680	ug/l	97
91) 1,2-Dichlorobenzene	14.082	146	337690	51.184	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.700	75	54794	49.649	ug/l	98
93) 1,2,4-Trichlorobenzene	15.364	180	185116	47.522	ug/l	99
94) Hexachlorobutadiene	15.470	225	64377	44.901	ug/l	97
95) Naphthalene	15.611	128	660404	52.814	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	187836	49.600	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
Data File : VN088013.D
Acq On : 03 Oct 2025 10:17
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025

Quant Time: Oct 04 02:51:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100325\
Data File : VN088013.D
Acq On : 03 Oct 2025 10:17
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025

Quant Time: Oct 04 02:51:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N092325W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 05:05:30 2025
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.206 min Scan# 1063

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:168 Resp: 211900

Ion Ratio Lower Upper

168 100





99 63.0 52.2 78.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025

Supervised By :Semsettin Yesilyurt 10/06/2025



#2

Dichlorodifluoromethane

Concen: 48.543 ug/l

RT: 2.142 min Scan# 32

Delta R.T. -0.000 min

Lab File: VN088013.D




Acq: 03 Oct 2025 10:17


Tgt Ion: 85 Resp: 119124

Ion Ratio Lower Upper

85 100


87 32.9 16.9 50.6





#3
Chloromethane
Concen: 49.489 ug/l
RT: 2.383 min Scan# 73
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 50 Resp: 124349
Ion Ratio Lower Upper
50 100
52 33.1 26.6 40.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#4
Vinyl Chloride
Concen: 50.707 ug/l
RT: 2.536 min Scan# 99
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion: 62 Resp: 129985
Ion Ratio Lower Upper
62 100
64 33.2 25.2 37.8





#5
Bromomethane
Concen: 51.258 ug/l
RT: 2.971 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 94 Resp: 81460
Ion Ratio Lower Upper
94 100
96 103.5 76.8 115.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#6
Chloroethane
Concen: 56.742 ug/l
RT: 3.142 min Scan# 202
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 64 Resp: 94576
Ion Ratio Lower Upper
64 100
66 33.1 23.8 35.8





#7
Trichlorofluoromethane
Concen: 54.280 ug/l
RT: 3.512 min Scan# 206
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 226552
Ion Ratio Lower Upper
101 100
103 65.0 50.2 75.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#8
Diethyl Ether
Concen: 54.177 ug/l
RT: 3.965 min Scan# 342
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17

Tgt Ion: 74 Resp: 85426
Ion Ratio Lower Upper
74 100
45 93.5 45.4 136.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.539 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 11810
Ion Ratio Lower Upper
101 100
85 46.1 36.5 54.7
151 73.6 58.6 87.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#10
Methyl Iodide
Concen: 55.410 ug/l
RT: 4.577 min Scan# 446
Delta R.T. -0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion:142 Resp: 110011
Ion Ratio Lower Upper
142 100
127 48.5 35.3 52.9
141 14.3 11.8 17.8





#11
Tert butyl alcohol
Concen: 234.750 ug/l
RT: 5.518 min Scan# 606
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 59 Resp: 120578
Ion Ratio Lower Upper
59 100
57 11.1 8.8 13.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#12
1,1-Dichloroethene
Concen: 53.163 ug/l
RT: 4.341 min Scan# 406
Delta R.T. 0.012 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion: 96 Resp: 118331
Ion Ratio Lower Upper
96 100
61 162.5 119.2 178.8
98 60.7 47.5 71.3





#13
Acrolein
Concen: 153.827 ug/l
RT: 4.177 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 10138
Ion Ratio Lower Upper
56 100
55 73.3 59.3 88.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#14
Allyl chloride
Concen: 46.845 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion: 41 Resp: 169750
Ion Ratio Lower Upper
41 100
39 87.6 63.8 95.8
76 41.7 33.0 49.6





#15
Acrylonitrile
Concen: 278.787 ug/l
RT: 5.706 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 449175
Ion Ratio Lower Upper
53 100
52 81.9 66.6 99.8
51 36.7 29.0 43.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#16
Acetone
Concen: 242.815 ug/l
RT: 4.418 min Scan# 419
Delta R.T. -0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 43 Resp: 356272
Ion Ratio Lower Upper
43 100
58 31.8 24.1 36.1





#17
Carbon Disulfide
Concen: 46.810 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 76 Resp: 325539
Ion Ratio Lower Upper
76 100
78 8.3 6.7 10.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#18
Methyl Acetate
Concen: 65.840 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17



Tgt Ion: 43 Resp: 241763
Ion Ratio Lower Upper
43 100
74 24.3 19.7 29.5









#19
Methyl tert-butyl Ether
Concen: 56.830 ug/l
RT: 5.788 min Scan# 651
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 73 Resp: 494130
Ion Ratio Lower Upper
73 100
57 20.6 16.3 24.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#20
Methylene Chloride
Concen: 55.203 ug/l
RT: 5.265 min Scan# 563
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 84 Resp: 163946
Ion Ratio Lower Upper
84 100
49 111.0 92.1 138.1
51 34.7 27.7 41.5
86 59.0 49.6 74.4





#21
trans-1,2-Dichloroethene
Concen: 53.418 ug/l
RT: 5.777 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 145759
Ion Ratio Lower Upper
96 100
61 123.0 114.4 171.6
98 62.9 53.6 80.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#22
Diisopropyl ether
Concen: 59.764 ug/l
RT: 6.653 min Scan# 799
Delta R.T. -0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 45 Resp: 505565
Ion Ratio Lower Upper
45 100
43 61.6 49.9 74.9
87 25.9 25.2 37.8
59 11.2 9.6 14.4





#23
 Vinyl Acetate
 Concen: 303.155 ug/l
 RT: 6.588 min Scan# 788
 Delta R.T. -0.000 min
 Lab File: VN088013.D
 Acq: 03 Oct 2025 10:17




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 215926
 Ion Ratio Lower Upper
 43 100
 86 11.0 8.6 12.8

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
 Supervised By :Semsettin Yesilyurt 10/06/2025




#24
 1,1-Dichloroethane
 Concen: 56.439 ug/l
 RT: 6.553 min Scan# 782
 Delta R.T. -0.000 min
 Lab File: VN088013.D
 Acq: 03 Oct 2025 10:17




Tgt Ion: 63 Resp: 278930
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.9 11.5
 100 3.9 2.1 6.5





#25
2-Butanone
Concen: 267.143 ug/l
RT: 7.465 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 599104
Ion Ratio Lower Upper
43 100
72 27.6 21.5 32.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#26
2,2-Dichloropropane
Concen: 32.031 ug/l
RT: 7.471 min Scan# 938
Delta R.T. -0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion: 77 Resp: 140646
Ion Ratio Lower Upper
77 100
97 25.3 11.5 34.5





#27
 cis-1,2-Dichloroethene
 Concen: 55.591 ug/l
 RT: 7.471 min Scan# 911
 Delta R.T. -0.000 min
 Lab File: VN088013.D
 Acq: 03 Oct 2025 10:17




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 96 Resp: 180488
 Ion Ratio Lower Upper
 96 100
 61 131.6 0.0 283.0
 98 62.6 0.0 127.6




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
 Supervised By :Semsettin Yesilyurt 10/06/2025



#28
 Bromochloromethane
 Concen: 48.057 ug/l
 RT: 7.800 min Scan# 994
 Delta R.T. 0.006 min
 Lab File: VN088013.D
 Acq: 03 Oct 2025 10:17

Tgt Ion: 49 Resp: 129042
 Ion Ratio Lower Upper
 49 100
 129 1.3 0.0 5.4
 130 75.8 62.6 94.0





#29

Tetrahydrofuran

Concen: 281.618 ug/l

RT: 7.824 min Scan# 998

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 40440

Ion	Ratio	Lower	Upper
42	100		
72	44.9	36.5	54.7
71	42.6	33.8	50.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#30

Chloroform

Concen: 57.299 ug/l

RT: 7.947 min Scan# 1019




Delta R.T. -0.000 min


Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Tgt Ion: 83 Resp: 309633


Ion	Ratio	Lower	Upper
83	100		
85	67.0	50.9	76.3





#31
Cyclohexane
Concen: 49.875 ug/l
RT: 8.241 min Scan# 1069
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 19944
Ion Ratio Lower Upper
56 100
69 30.6 23.4 35.0
84 92.4 70.8 106.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#32
1,1,1-Trichloroethane
Concen: 55.224 ug/l
RT: 8.147 min Scan# 1053
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 97 Resp: 252160
Ion Ratio Lower Upper
97 100
99 67.5 51.1 76.7
61 49.3 38.6 58.0





#33
1,2-Dichloroethane-d4
Concen: 53.155 ug/l
RT: 8.559 min Scan# 1123
Delta R.T. -0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 65 Resp: 189558
Ion Ratio Lower Upper
65 100
67 52.7 0.0 103.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.082 min Scan# 1212
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17



Tgt Ion:114 Resp: 407793
Ion Ratio Lower Upper
114 100
63 20.6 0.0 40.0
88 15.1 0.0 31.8





#35

Dibromofluoromethane

Concen: 49.840 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. -0.000 min

Lab File: VN088013.D


Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050






Tgt Ion: 113 Resp: 147560

Ion	Ratio	Lower	Upper
113	100		
111	101.6	82.2	123.2
192	17.2	14.2	21.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#36

1,1-Dichloropropene

Concen: 51.838 ug/l

RT: 8.353 min Scan# 1088




Delta R.T. -0.000 min


Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Tgt Ion: 75 Resp: 184552

Ion	Ratio	Lower	Upper
75	100		
110	35.4	17.1	51.3
77	32.5	24.8	37.2





#37

Ethyl Acetate

Concen: 54.045 ug/l

RT: 7.547 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 25266

Ion Ratio Lower Upper

43 100

61 13.4 10.7 16.1





70 10.0 7.8 11.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025

Supervised By :Semsettin Yesilyurt 10/06/2025



#38

Carbon Tetrachloride

Concen: 52.458 ug/l

RT: 8.347 min Scan# 1087

Delta R.T. 0.006 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17




Tgt Ion:117 Resp: 214305


Ion Ratio Lower Upper

117 100

119 93.9 72.1 108.1


121 33.6 24.6 36.8





#39
Methylcyclohexane
Concen: 47.480 ug/l
RT: 9.582 min Scan# 1129
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 186475
Ion Ratio Lower Upper
83 100
55 80.5 59.9 89.9
98 52.6 37.4 56.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#40
Benzene
Concen: 53.533 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 78 Resp: 631046
Ion Ratio Lower Upper
78 100
77 24.9 19.0 28.6





#41
Methacrylonitrile
Concen: 52.700 ug/l
RT: 7.759 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 127328
Ion Ratio Lower Upper
41 100
39 56.6 43.5 65.3
67 76.0 57.6 86.4
52 34.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#42
1,2-Dichloroethane
Concen: 52.013 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 62 Resp: 228809
Ion Ratio Lower Upper
62 100
98 9.7 0.0 18.6





#43
Isopropyl Acetate
Concen: 54.310 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 39966
Ion Ratio Lower Upper
43 100
61 25.2 20.2 30.2
87 12.9 10.9 16.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#44
Trichloroethene
Concen: 50.528 ug/l
RT: 9.329 min Scan# 1254
Delta R.T. -0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17

Tgt Ion:130 Resp: 142273
Ion Ratio Lower Upper
130 100
95 97.6 0.0 208.0





#45

1,2-Dichloropropane

Concen: 54.312 ug/l

RT: 9.600 min Scan# 1300

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 160540

Ion Ratio Lower Upper

63 100





65 30.6 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025

Supervised By :Semsettin Yesilyurt 10/06/2025



#46

Dibromomethane

Concen: 55.256 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17




Tgt Ion: 93 Resp: 128880


Ion Ratio Lower Upper

93 100

95 81.6 67.6 101.4

174 84.7 75.4 113.0









#47
Bromodichloromethane
Concen: 55.172 ug/l
RT: 9.871 min Scan# 1346
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 25239
Ion Ratio Lower Upper
83 100
85 67.5 49.1 73.7
127 6.5 5.9 8.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#48
Methyl methacrylate
Concen: 55.005 ug/l
RT: 9.665 min Scan# 1311
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 41 Resp: 187451
Ion Ratio Lower Upper
41 100
69 84.5 70.9 106.3
39 64.8 52.9 79.3





#49
1,4-Dioxane
Concen: 904.603 ug/l
RT: 9.682 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 88 Resp: 4554
Ion Ratio Lower Upper
88 100
43 0.0 22.6 33.8
58 72.5 57.4 86.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#50
Toluene-d8
Concen: 49.548 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion: 98 Resp: 515889
Ion Ratio Lower Upper
98 100
100 63.9 51.7 77.5





#51
4-Methyl-2-Pentanone
Concen: 282.622 ug/l
RT: 10.423 min Scan# 1440
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 129778
Ion Ratio Lower Upper
43 100
58 38.5 31.0 46.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#52
Toluene
Concen: 52.688 ug/l
RT: 10.606 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 92 Resp: 389751
Ion Ratio Lower Upper
92 100
91 178.6 138.2 207.2





#53
t-1,3-Dichloropropene
Concen: 49.522 ug/l
RT: 10.818 min Scan# 1507
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 23186
Ion Ratio Lower Upper
75 100
77 32.2 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#54
cis-1,3-Dichloropropene
Concen: 48.009 ug/l
RT: 10.294 min Scan# 1418
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion: 75 Resp: 233459
Ion Ratio Lower Upper
75 100
77 31.9 24.7 37.1
39 52.8 41.4 62.0





#55
1,1,2-Trichloroethane
Concen: 54.906 ug/l
RT: 10.994 min Scan# 1537
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 97 Resp: 16514
Ion Ratio Lower Upper
97 100
83 91.6 69.0 103.4
85 57.2 44.4 66.6
99 64.5 52.3 78.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#56
Ethyl methacrylate
Concen: 57.638 ug/l
RT: 10.853 min Scan# 1513
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17



Tgt Ion: 69 Resp: 250497
Ion Ratio Lower Upper
69 100
41 65.0 53.8 80.6
39 41.2 34.7 52.1





#57

1,3-Dichloropropane

Concen: 54.717 ug/l

RT: 11.141 min Scan# 1562

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 270690

Ion Ratio Lower Upper





76 100

78 32.5 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 355.558 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. -0.000 min

Lab File: VN088013.D




Acq: 03 Oct 2025 10:17


Tgt Ion: 63 Resp: 726559

Ion Ratio Lower Upper

63 100


106 28.3 22.6 33.8





#59
2-Hexanone
Concen: 289.453 ug/l
RT: 11.176 min Scan# 1568
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 878869
Ion Ratio Lower Upper
43 100
58 53.4 26.4 79.2

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#60
Dibromochloromethane
Concen: 54.865 ug/l
RT: 11.341 min Scan# 1596
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion:129 Resp: 194024
Ion Ratio Lower Upper
129 100
127 77.5 39.3 117.9





#61
1,2-Dibromoethane
Concen: 54.463 ug/l
RT: 11.447 min Scan# 1614
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:107 Resp: 173418
Ion Ratio Lower Upper
107 100
109 93.3 75.3 112.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#62
4-Bromofluorobenzene
Concen: 48.498 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion: 95 Resp: 180495
Ion Ratio Lower Upper
95 100
174 79.0 0.0 159.2
176 74.4 0.0 152.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.847 min Scan# 1681
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:117 Resp: 399348
Ion Ratio Lower Upper
117 100
82 53.8 47.0 70.4
119 30.5 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#64
Tetrachloroethene
Concen: 52.158 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion:164 Resp: 126880
Ion Ratio Lower Upper
164 100
166 128.6 101.6 152.4
129 101.4 77.5 116.3
131 90.1 73.4 110.0





#65
Chlorobenzene
Concen: 51.989 ug/l
RT: 11.870 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:112 Resp: 441827
Ion Ratio Lower Upper
112 100
114 33.9 27.2 40.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 53.146 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17



Tgt Ion:131 Resp: 162876
Ion Ratio Lower Upper
131 100
133 93.8 48.4 145.0
119 67.2 31.9 95.5





#67

Ethyl Benzene

Concen: 52.535 ug/l

RT: 11.941 min Scan# 1698

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 744970

Ion Ratio Lower Upper

91 100





106 30.1 25.2 37.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025

Supervised By :Semsettin Yesilyurt 10/06/2025



#68

m/p-Xylenes

Concen: 106.758 ug/l

RT: 12.047 min Scan# 1716

Delta R.T. -0.000 min

Lab File: VN088013.D




Acq: 03 Oct 2025 10:17


Tgt Ion:106 Resp: 578990

Ion Ratio Lower Upper

106 100

91 206.6 162.9 244.3









#69
o-Xylene
Concen: 53.589 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 279227
Ion Ratio Lower Upper
106 100
91 214.6 108.2 324.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#70
Styrene
Concen: 54.825 ug/l
RT: 12.388 min Scan# 1774
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion:104 Resp: 490624
Ion Ratio Lower Upper
104 100
78 50.3 39.7 59.5
103 55.3 44.6 67.0





#71
Bromoform
Concen: 53.359 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 132670
Ion Ratio Lower Upper
173 100
175 47.1 23.7 71.1
254 0.1 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17



Tgt Ion:152 Resp: 212700
Ion Ratio Lower Upper
152 100
115 61.6 29.9 89.7
150 178.0 0.0 335.0





#73

Isopropylbenzene

Concen: 53.910 ug/l

RT: 12.676 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 71067

Ion Ratio Lower Upper

105 100





120 26.3 13.1 39.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025

Supervised By :Semsettin Yesilyurt 10/06/2025



#74

N-ethyl acetate

Concen: 51.938 ug/l m

RT: 12.506 min Scan# 1794

Delta R.T. 0.006 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Tgt Ion: 43 Resp: 225597




Ion Ratio Lower Upper


43 100

70 18.3 19.8 29.8#

55 12.5 9.6 14.4


61 12.2 7.8 11.6#





#75
 1,1,2,2-Tetrachloroethane
 Concen: 53.479 ug/l
 RT: 12.917 min Scan# 1864
 Delta R.T. 0.006 min
 Lab File: VN088013.D
 Acq: 03 Oct 2025 10:17




Instrument : MSVOA_N
 ClientSampleId : VSTDCCC050



Tgt Ion: 83 Resp: 254009

Ion	Ratio	Lower	Upper
83	100		
131	9.2	5.2	15.6
85	65.1	32.4	97.0




Manual Integrations
APPROVED
 Reviewed By :Mahesh Dadoda 10/06/2025
 Supervised By :Semsettin Yesilyurt 10/06/2025




#76
 1,2,3-Trichloropropane
 Concen: 47.218 ug/l m
 RT: 12.970 min Scan# 1873
 Delta R.T. -0.000 min
 Lab File: VN088013.D
 Acq: 03 Oct 2025 10:17

Tgt Ion: 75 Resp: 197976


Ion	Ratio	Lower	Upper
75	100		
77	202.2	89.5	268.5





#77
Bromobenzene
Concen: 52.440 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:156 Resp: 183930
Ion Ratio Lower Upper
156 100
77 217.7 108.2 324.6
158 95.8 49.6 149.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#78
n-propylbenzene
Concen: 52.889 ug/l
RT: 13.011 min Scan# 1880
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion: 91 Resp: 871697
Ion Ratio Lower Upper
91 100
120 22.0 11.1 33.1





#79
2-Chlorotoluene
Concen: 51.754 ug/l
RT: 13.100 min Scan# 1895
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 51938
Ion Ratio Lower Upper
91 100
126 33.8 16.6 49.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#80
1,3,5-Trimethylbenzene
Concen: 53.110 ug/l
RT: 13.147 min Scan# 1903
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion:105 Resp: 605591
Ion Ratio Lower Upper
105 100
120 48.2 24.7 74.1





#81
trans-1,4-Dichloro-2-butene
Concen: 43.414 ug/l
RT: 12.717 min Scan# 1830
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 75 Resp: 65552
Ion Ratio Lower Upper
75 100
53 104.1 82.0 123.0
89 47.3 36.1 54.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#82
4-Chlorotoluene
Concen: 53.641 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion: 91 Resp: 542423
Ion Ratio Lower Upper
91 100
126 33.5 16.9 50.6





#83
tert-Butylbenzene
Concen: 54.079 ug/l
RT: 13.417 min Scan# 1948
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:119 Resp: 52194
Ion Ratio Lower Upper
119 100
91 65.4 32.1 96.3
134 27.2 12.3 36.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 53.814 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17

Tgt Ion:105 Resp: 629394
Ion Ratio Lower Upper
105 100
120 45.4 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 52.962 ug/l
 RT: 13.594 min Scan# 1979
 Delta R.T. -0.000 min
 Lab File: VN088013.D
 Acq: 03 Oct 2025 10:17

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Tgt Ion:105 Resp: 749419
 Ion Ratio Lower Upper
 105 100
 134 20.3 9.6 28.8




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
 Supervised By :Semsettin Yesilyurt 10/06/2025



#86
 p-Isopropyltoluene
 Concen: 52.499 ug/l
 RT: 13.706 min Scan# 1998
 Delta R.T. -0.000 min
 Lab File: VN088013.D
 Acq: 03 Oct 2025 10:17

Tgt Ion:119 Resp: 629181
 Ion Ratio Lower Upper
 119 100
 134 25.9 13.1 39.3
 91 24.6 11.9 35.5





#87

1,3-Dichlorobenzene

Concen: 52.270 ug/l

RT: 13.711 min Scan# 1999

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:146 Resp: 35776

Ion Ratio Lower Upper

146 100

111 41.2 20.3 60.8





148 63.6 32.4 97.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025

Supervised By :Semsettin Yesilyurt 10/06/2025



#88

1,4-Dichlorobenzene

Concen: 52.078 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17




Tgt Ion:146 Resp: 366323


Ion Ratio Lower Upper

146 100

111 39.4 20.5 61.5

148 62.5 32.8 98.4









#89
n-Butylbenzene
Concen: 49.962 ug/l
RT: 14.035 min Scan# 2053
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 56587
Ion Ratio Lower Upper
91 100
92 51.8 25.9 77.8
134 25.1 12.6 37.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#90
Hexachloroethane
Concen: 51.680 ug/l
RT: 14.311 min Scan# 2101
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17


Tgt Ion:117 Resp: 129377
Ion Ratio Lower Upper
117 100
201 64.8 33.6 100.6





#91
1,2-Dichlorobenzene
Concen: 51.184 ug/l
RT: 14.082 min Scan# 2062
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 337690
Ion Ratio Lower Upper
146 100
111 42.3 20.9 62.7
148 65.2 31.9 95.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 49.649 ug/l
RT: 14.700 min Scan# 2167
Delta R.T. 0.006 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Tgt Ion: 75 Resp: 54794
Ion Ratio Lower Upper
75 100
155 78.9 40.2 120.6
157 98.7 50.9 152.7





#93
1,2,4-Trichlorobenzene
Concen: 47.522 ug/l
RT: 15.364 min Scan# 2280
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:180 Resp: 185110
Ion Ratio Lower Upper
180 100
182 96.7 48.5 145.5
145 32.5 15.9 47.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025




#94
Hexachlorobutadiene
Concen: 44.901 ug/l
RT: 15.470 min Scan# 2298
Delta R.T. -0.000 min
Lab File: VN088013.D
Acq: 03 Oct 2025 10:17



Tgt Ion:225 Resp: 64377
Ion Ratio Lower Upper
225 100
223 65.6 31.8 95.4
227 66.0 31.9 95.9





#95

Naphthalene

Concen: 52.814 ug/l

RT: 15.611 min Scan# 2322

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion:	128	Resp:	66040
Ion Ratio	Lower	Upper	
128	100		
127	13.3	10.2	15.4
129	10.6	8.6	13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/06/2025
Supervised By :Semsettin Yesilyurt 10/06/2025



#96

1,2,3-Trichlorobenzene

Concen: 49.600 ug/l

RT: 15.811 min Scan# 2356

Delta R.T. -0.000 min

Lab File: VN088013.D

Acq: 03 Oct 2025 10:17

Tgt Ion:	180	Resp:	187836
Ion Ratio	Lower	Upper	
180	100		
182	96.6	47.3	142.0
145	35.1	16.1	48.3

