

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100418\
 Data File : VN051671.D
 Acq On : 4 Oct 2018 18:19
 Operator : MD\SY
 Sample : PB113487ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113487ZHE#03

Quant Time: Oct 05 07:09:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 05 02:22:33 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	639345	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	942127	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	828816	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	318348	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	400024	47.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.72%	
35) Dibromofluoromethane	7.59	113	355935	48.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.74%	
50) Toluene-d8	10.09	98	1356487	49.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.84%	
62) 4-Bromofluorobenzene	12.40	95	386492	40.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	80.36%	
Target Compounds						
20) Methylene Chloride	4.56	84	11474	1.651	ug/l	92
43) Isopropyl Acetate	8.17	43	320505	29.931	ug/l #	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100418\
Data File : VN051671.D
Acq On : 4 Oct 2018 18:19
Operator : MD\SY
Sample : PB113487ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113487ZHE#03

Quant Time: Oct 05 07:09:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N100418W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 05 02:22:33 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1834

Delta R.T. 0.00 min

Lab File: VN051671.D

Acq: 4 Oct 2018 18:19

Instrument :

MSVOA_N

ClientSampleId :





PB113487ZHE#03

Tot Ion:168 Resp: 639345

Ion Ratio Lower Upper

168 100

99 48.7 40.4 60.6



#20

Methylene Chloride

Concen: 1.651 ug/l

RT: 4.56 min Scan# 868

Delta R.T. 0.01 min

Lab File: VN051671.D

Acq: 4 Oct 2018 18:19

Tgt Ion: 84 Resp: 11474




Ion Ratio Lower Upper


84 100

49 98.3 89.5 134.3

51 38.0 27.4 41.0

86 65.1 51.9 77.9









#33
 1,2-Dichloroethane-d4
 Concen: Below ug/l
 RT: 8.03 min Scan# 1946
 Delta R.T. -0.00 min
 Lab File: VN051671.D
 Acq: 4 Oct 2018 18:19




Instrument :
 MSVOA_N
 ClientSampleId :
 PB113487ZHE#03


Tot Ion: 65 Resp: 400024
 Ion Ratio Lower Upper
 65 100
 67 52.3 0.0 103.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. -0.00 min
 Lab File: VN051671.D
 Acq: 4 Oct 2018 18:19

Tgt Ion: 114 Resp: 942127
 Ion Ratio Lower Upper
 114 100
 63 18.5 0.0 36.2
 88 14.6 0.0 30.2






#35
Dibromofluoromethane
Concen: 50.000 ug/l
RT: 7.59 min Scan# 1810
Delta R.T. -0.00 min
Lab File: VN051671.D
Acq: 4 Oct 2018 18:19

Instrument :
MSVOA_N
ClientSampleId :
PB113487ZHE#03

Tot Ion	Ratio	Lower	Upper
113	100		
111	103.4	81.8	122.8
192	20.9	16.5	24.7






Abundance

Ion 112.80 (112.50 to 113.50): V


Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#43
Isopropyl Acetate
Concen: 29.931 ug/l
RT: 8.17 min Scan# 1989
Delta R.T. 0.00 min
Lab File: VN051671.D
Acq: 4 Oct 2018 18:19

Tgt Ion	Ratio	Lower	Upper
43	100		
61	16.6	15.8	23.6
87	12.6	12.6	19.0






Abundance

Ion 43.00 (42.70 to 43.70): V

Ion 61.00 (60.70 to 61.70): V

Ion 87.00 (86.70 to 87.70): V





#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN051671.D

Acq: 4 Oct 2018 18:19

Instrument :

MSVOA_N

ClientSampleId :





PB113487ZHE#03

Tot Ion: 98 Resp: 1356487

Ion Ratio Lower Upper

98 100

100 63.2 51.5 77.3



#62

4-Bromofluorobenzene

Concen: Below ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN051671.D

Acq: 4 Oct 2018 18:19




Tgt Ion: 95 Resp: 386492


Ion Ratio Lower Upper

95 100

174 94.3 0.0 182.8

176 90.5 0.0 176.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2997
Delta R.T. 0.00 min
Lab File: VN051671.D
Acq: 4 Oct 2018 18:19

Instrument :
MSVOA_N
ClientSampleId :
PB113487ZHE#03

Tot Ion:117 Resp: 828816
Ion Ratio Lower Upper
117 100
82 51.8 41.3 61.9
119 32.4 25.9 38.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN051671.D
Acq: 4 Oct 2018 18:19

Tgt Ion:152 Resp: 318348
Ion Ratio Lower Upper
152 100
115 55.6 28.6 85.8
150 154.4 0.0 352.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

