

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100419\
 Data File : VN058531.D
 Acq On : 4 Oct 2019 11:47
 Operator : JC/SP
 Sample : K5169-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-100219

Quant Time: Oct 05 00:31:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	413079	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	683834	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	596683	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	255249	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	299419	49.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.20%	
35) Dibromofluoromethane	7.57	113	207198	49.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.88%	
50) Toluene-d8	10.08	98	847435	51.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.22%	
62) 4-Bromofluorobenzene	12.41	95	285787	48.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100419\
Data File : VN058531.D
Acq On : 4 Oct 2019 11:47
Operator : JC/SP
Sample : K5169-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-100219

Quant Time: Oct 05 00:31:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

