

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100419\
 Data File : VN058542.D
 Acq On : 4 Oct 2019 15:47
 Operator : JC/SP
 Sample : K5169-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-20D_100119

Quant Time: Oct 16 01:37:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	414536	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	667883	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	584783	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	242752	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	296004	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
35) Dibromofluoromethane	7.57	113	209181	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
50) Toluene-d8	10.08	98	837691	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
62) 4-Bromofluorobenzene	12.41	95	285594	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
Target Compounds						
3) Chloromethane	2.04	50	2263	0.453	ug/l #	85
12) 1,1-Dichloroethene	3.72	96	5253	1.343	ug/l	95
16) Acetone	3.80	43	8719	4.222	ug/l #	87
19) Methyl tert-butyl Ether	5.02	73	38665	2.802	ug/l	98
27) cis-1,2-Dichloroethene	6.81	96	24616	4.843	ug/l	87
44) Trichloroethene	8.82	130	60927	13.442	ug/l	94
64) Tetrachloroethene	10.63	164	53241	14.330	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100419\
Data File : VN058542.D
Acq On : 4 Oct 2019 15:47
Operator : JC/SP
Sample : K5169-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-20D_100119

Quant Time: Oct 16 01:37:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

