

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100419\
 Data File : VN058543.D
 Acq On : 4 Oct 2019 16:09
 Operator : JC/SP
 Sample : K5169-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-9R 100119

Quant Time: Oct 05 03:31:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	425183	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	687405	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	593598	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	240949	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	296388	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
35) Dibromofluoromethane	7.57	113	208609	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
50) Toluene-d8	10.08	98	854228	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
62) 4-Bromofluorobenzene	12.41	95	279114	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
Target Compounds						
12) 1,1-Dichloroethene	3.72	96	9407	2.345	ug/l	91
16) Acetone	3.80	43	9042	4.269	ug/l	96
27) cis-1,2-Dichloroethene	6.81	96	48185	9.243	ug/l	85
29) Tetrahydrofuran	7.21	42	2306	1.249	ug/l #	57
30) Chloroform	7.36	83	3705	0.414	ug/l	88
32) 1,1,1-Trichloroethane	7.55	97	26999	3.670	ug/l	98
44) Trichloroethene	8.83	130	53858	11.545	ug/l	98
64) Tetrachloroethene	10.62	164	7870	2.087	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100419\
Data File : VN058543.D
Acq On : 4 Oct 2019 16:09
Operator : JC/SP
Sample : K5169-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-9R_100119

Quant Time: Oct 05 03:31:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

