

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100419\
 Data File : VN058549.D
 Acq On : 4 Oct 2019 18:21
 Operator : JC/SP
 Sample : K5197-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-8 100219

Quant Time: Oct 05 04:26:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	424148	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	681841	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	591812	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	237251	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	294832	47.57	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.14%	
35) Dibromofluoromethane	7.57	113	208841	49.97	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.94%	
50) Toluene-d8	10.09	98	875900	53.50	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	107.00%	
62) 4-Bromofluorobenzene	12.41	95	274240	46.33	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.66%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.04	50	2345	0.459	ug/l	95
4) Vinyl Chloride	2.17	62	30928	6.175	ug/l	99
16) Acetone	3.80	43	6937	3.283	ug/l	94
21) trans-1,2-Dichloroethene	5.01	96	1545	0.347	ug/l #	66
27) cis-1,2-Dichloroethene	6.81	96	219986	42.303	ug/l	89
44) Trichloroethene	8.82	130	59226	12.799	ug/l	91
64) Tetrachloroethene	10.63	164	244101	64.921	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100419\
Data File : VN058549.D
Acq On : 4 Oct 2019 18:21
Operator : JC/SP
Sample : K5197-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-8 100219

Quant Time: Oct 05 04:26:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

