

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100419\
 Data File : VN058552.D
 Acq On : 4 Oct 2019 19:28
 Operator : JC/SP
 Sample : K5193-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20190880-COMPOSITE-L

Quant Time: Oct 05 04:53:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	411262	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	659585	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	592931	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	240735	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	296749	49.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.76%	
35) Dibromofluoromethane	7.57	113	188086	46.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.06%	
50) Toluene-d8	10.08	98	839439	53.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.02%	
62) 4-Bromofluorobenzene	12.41	95	276431	48.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.56%	
Target Compounds						
16) Acetone	3.80	43	61439	29.989	ug/l	100
17) Carbon Disulfide	4.03	76	23135	2.262	ug/l #	92
20) Methylene Chloride	4.53	84	1340107	283.446	ug/l	99
25) 2-Butanone	6.82	43	14879	5.290	ug/l	96
43) Isopropyl Acetate	8.15	43	10311	1.025	ug/l	98
65) Chlorobenzene	11.43	112	32512	2.759	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100419\
Data File : VN058552.D
Acq On : 4 Oct 2019 19:28
Operator : JC/SP
Sample : K5193-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20190880-COMPOSITE-L

Quant Time: Oct 05 04:53:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

