

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100419\
 Data File : VN058555.D
 Acq On : 4 Oct 2019 20:34
 Operator : JC/SP
 Sample : K5197-10 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-23S_100319

Quant Time: Oct 05 05:05:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 30 08:07:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	421416	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	674133	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	596803	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	250321	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	301649	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
35) Dibromofluoromethane	7.57	113	210330	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
50) Toluene-d8	10.08	98	871197	53.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.66%	
62) 4-Bromofluorobenzene	12.41	95	288506	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
Target Compounds						
3) Chloromethane	2.04	50	1789	0.352	ug/l #	85
27) cis-1,2-Dichloroethene	6.81	96	4547	0.880	ug/l	95
44) Trichloroethene	8.82	130	39336	8.598	ug/l	95
64) Tetrachloroethene	10.63	164	1494445	394.138	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100419\
Data File : VN058555.D
Acq On : 4 Oct 2019 20:34
Operator : JC/SP
Sample : K5197-10 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-23S_100319

Quant Time: Oct 05 05:05:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092719W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 30 08:07:38 2019
Response via : Initial Calibration

