

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079404.D
 Acq On : 04 Oct 2023 15:35
 Operator : JC\MD
 Sample : 04689-01 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-1

Quant Time: Oct 05 02:22:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	146688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	272273	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	238681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	66427	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	104970	53.011	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.020%
35) Dibromofluoromethane	8.171	113	81839	45.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.380%
50) Toluene-d8	10.571	98	251406	37.805	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	75.600%#
62) 4-Bromofluorobenzene	12.853	95	81885	38.602	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	77.200%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079404.D
Acq On : 04 Oct 2023 15:35
Operator : JC\MD
Sample : 04689-01 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-1

Quant Time: Oct 05 02:22:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

