

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079405.D
 Acq On : 04 Oct 2023 15:59
 Operator : JC\MD
 Sample : PB156087ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156087ZHE#03

Quant Time: Oct 05 02:22:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	122350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	247447	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	248805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	60528	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	109847	66.509	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 133.020%#	
35) Dibromofluoromethane	8.171	113	92836	56.402	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 112.800%	
50) Toluene-d8	10.571	98	338244	55.966	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.940%	
62) 4-Bromofluorobenzene	12.853	95	88370	45.839	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.680%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	22716	32.384	ug/l	91
20) Methylene Chloride	5.277	84	15878	9.302	ug/l #	81
43) Isopropyl Acetate	8.694	43	135447	32.804	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079405.D
Acq On : 04 Oct 2023 15:59
Operator : JC\MD
Sample : PB156087ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156087ZHE#03

Quant Time: Oct 05 02:22:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

