

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079406.D
 Acq On : 04 Oct 2023 16:23
 Operator : JC\MD
 Sample : PB156087ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156087ZHE#04

Quant Time: Oct 05 02:23:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	127471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	262315	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	273857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	71204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	114685	66.649	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 133.300%#			
35) Dibromofluoromethane	8.171	113	95315	54.626	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.260%			
50) Toluene-d8	10.571	98	355038	55.415	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.820%			
62) 4-Bromofluorobenzene	12.853	95	97133	47.529	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.060%			
Target Compounds						
16) Acetone	4.442	43	22112	30.257	ug/l	97
20) Methylene Chloride	5.277	84	16056	9.028	ug/l #	74
43) Isopropyl Acetate	8.694	43	136210	31.119	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079406.D
Acq On : 04 Oct 2023 16:23
Operator : JC\MD
Sample : PB156087ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156087ZHE#04

Quant Time: Oct 05 02:23:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

