

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079408.D
 Acq On : 04 Oct 2023 17:11
 Operator : JC\MD
 Sample : PB156087ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156087ZHE#06

Quant Time: Oct 05 02:23:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	123759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	257256	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	261318	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	65968	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	113996	68.235	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 136.480%#	
35) Dibromofluoromethane	8.171	113	93607	54.702	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.400%	
50) Toluene-d8	10.570	98	342133	54.451	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.900%	
62) 4-Bromofluorobenzene	12.853	95	93914	46.857	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.720%	
Target Compounds						
						Qvalue
16) Acetone	4.441	43	24104	33.972	ug/l	91
20) Methylene Chloride	5.277	84	25724	14.898	ug/l #	87
43) Isopropyl Acetate	8.694	43	151239	35.232	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079408.D
Acq On : 04 Oct 2023 17:11
Operator : JC\MD
Sample : PB156087ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156087ZHE#06

Quant Time: Oct 05 02:23:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

