

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079410.D
 Acq On : 04 Oct 2023 18:00
 Operator : JC\MD
 Sample : 04712-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Oct 05 02:24:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	128122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	258959	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	239224	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	60397	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	117471	67.921	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	135.840%#
35) Dibromofluoromethane	8.171	113	92612	53.765	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.520%
50) Toluene-d8	10.571	98	302541	47.833	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.660%
62) 4-Bromofluorobenzene	12.853	95	92957	46.075	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.140%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	28235	38.439	ug/l	91
20) Methylene Chloride	5.277	84	33323	18.642	ug/l #	78
43) Isopropyl Acetate	8.694	43	182404	42.212	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079410.D
Acq On : 04 Oct 2023 18:00
Operator : JC\MD
Sample : 04712-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Oct 05 02:24:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

