

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079411.D
 Acq On : 04 Oct 2023 18:24
 Operator : JC\MD
 Sample : 04712-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2

Quant Time: Oct 05 02:25:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	129935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	264398	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	241962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	62348	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	118777	67.718	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 135.440%#	
35) Dibromofluoromethane	8.177	113	89794	51.056	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.120%	
50) Toluene-d8	10.571	98	298958	46.294	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 92.580%	
62) 4-Bromofluorobenzene	12.853	95	95061	46.148	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.300%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	27741	37.240	ug/l #	83
18) Methyl Acetate	5.042	43	4137	1.351	ug/l #	60
20) Methylene Chloride	5.283	84	33107	18.263	ug/l #	78
43) Isopropyl Acetate	8.700	43	185285	41.997	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079411.D
Acq On : 04 Oct 2023 18:24
Operator : JC\MD
Sample : 04712-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2

Quant Time: Oct 05 02:25:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

