

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079412.D
 Acq On : 04 Oct 2023 18:48
 Operator : JC\MD
 Sample : 04714-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Oct 05 02:25:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	126388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	255050	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	230532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	58845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	118011	69.169	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 138.340%#	
35) Dibromofluoromethane	8.171	113	88147	51.957	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.920%	
50) Toluene-d8	10.570	98	291676	46.822	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 93.640%	
62) 4-Bromofluorobenzene	12.853	95	88991	44.785	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.580%	
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	33800	46.647	ug/l #	87
20) Methylene Chloride	5.277	84	33759	19.145	ug/l	87
25) 2-Butanone	7.494	43	5718	5.439	ug/l #	76
43) Isopropyl Acetate	8.694	43	185147	43.504	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079412.D
Acq On : 04 Oct 2023 18:48
Operator : JC\MD
Sample : 04714-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Oct 05 02:25:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

