

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079413.D
 Acq On : 04 Oct 2023 19:12
 Operator : JC\MD
 Sample : 04714-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Oct 05 02:26:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	127917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	262860	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	239540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	59736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	117407	67.993	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 135.980%#	
35) Dibromofluoromethane	8.171	113	91177	52.146	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.300%	
50) Toluene-d8	10.571	98	298669	46.520	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 93.040%	
62) 4-Bromofluorobenzene	12.853	95	89690	43.796	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.600%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	46512	63.423	ug/l	94
18) Methyl Acetate	5.047	43	5735	1.903	ug/l #	81
20) Methylene Chloride	5.277	84	26683	14.951	ug/l #	83
25) 2-Butanone	7.506	43	11047	10.382	ug/l #	74
43) Isopropyl Acetate	8.694	43	146195	33.331	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079413.D
Acq On : 04 Oct 2023 19:12
Operator : JC\MD
Sample : 04714-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Oct 05 02:26:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

