

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079414.D
 Acq On : 04 Oct 2023 19:36
 Operator : JC\MD
 Sample : 04714-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Oct 05 02:26:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	120034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	243948	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	221711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	57510	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	113286	69.915	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 139.820%#	
35) Dibromofluoromethane	8.171	113	88683	54.652	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.300%	
50) Toluene-d8	10.571	98	298113	50.033	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.060%	
62) 4-Bromofluorobenzene	12.853	95	93379	49.132	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.260%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	42996	62.479	ug/l	96
20) Methylene Chloride	5.289	84	31989	19.102	ug/l	96
25) 2-Butanone	7.489	43	9926	9.941	ug/l #	63
43) Isopropyl Acetate	8.694	43	167025	41.032	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079414.D
Acq On : 04 Oct 2023 19:36
Operator : JC\MD
Sample : 04714-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Oct 05 02:26:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

