

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
 Data File : VN079400.D
 Acq On : 04 Oct 2023 13:38
 Operator : JC\MD
 Sample : VN1004WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1004WBL01

Quant Time: Oct 05 02:19:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	144184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	269515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	233046	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	69824	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	115561	59.373	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.740%
35) Dibromofluoromethane	8.171	113	91423	50.996	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.000%
50) Toluene-d8	10.571	98	288251	43.789	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.580%
62) 4-Bromofluorobenzene	12.853	95	88924	42.349	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.700%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN100423\
Data File : VN079400.D
Acq On : 04 Oct 2023 13:38
Operator : JC\MD
Sample : VN1004WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1004WBL01

Quant Time: Oct 05 02:19:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1067
Delta R.T. -0.000 min
Lab File: VN079400.D
Acq: 04 Oct 2023 13:38




Instrument :
MSVOA_N
ClientSampleId :
VN1004WBL01


Tgt Ion:168 Resp: 144184
Ion Ratio Lower Upper
168 100
99 60.3 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 59.373 ug/l
RT: 8.582 min Scan# 1127
Delta R.T. -0.001 min
Lab File: VN079400.D
Acq: 04 Oct 2023 13:38

Tgt Ion: 65 Resp: 115561
Ion Ratio Lower Upper
65 100
67 51.4 0.0 106.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN079400.D

Acq: 04 Oct 2023 13:38

Instrument :

MSVOA_N

ClientSampleId :

VN1004WBL01





Tgt Ion:114 Resp: 269515

Ion Ratio Lower Upper

114 100

63 23.0 0.0 36.4

88 17.4 0.0 30.6



#35

Dibromofluoromethane

Concen: 50.996 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN079400.D

Acq: 04 Oct 2023 13:38




Tgt Ion:113 Resp: 91423


Ion Ratio Lower Upper

113 100

111 103.3 82.6 123.8

192 17.3 12.7 19.1









#50
Toluene-d8
Concen: 43.789 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. -0.000 min
Lab File: VN079400.D
Acq: 04 Oct 2023 13:38




Instrument :
MSVOA_N
ClientSampleId :
VN1004WBL01


Tgt Ion: 98 Resp: 288251
Ion Ratio Lower Upper
98 100
100 64.0 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 42.349 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VN079400.D
Acq: 04 Oct 2023 13:38

Tgt Ion: 95 Resp: 88924
Ion Ratio Lower Upper
95 100
174 76.2 0.0 143.8
176 73.9 0.0 135.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.871 min Scan# 1

Delta R.T. -0.001 min

Lab File: VN079400.D

Acq: 04 Oct 2023 13:38

Instrument :

MSVOA_N

ClientSampleId :

VN1004WBL01





Tgt Ion:117 Resp: 233046

Ion Ratio Lower Upper

117 100

82 57.6 44.0 66.0

119 31.3 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. -0.000 min

Lab File: VN079400.D

Acq: 04 Oct 2023 13:38

Tgt Ion:152 Resp: 69824

Ion Ratio Lower Upper

152 100

115 60.9 32.3 96.8

150 155.5 0.0 355.8

