

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN100518\
 Data File : VN051684.D
 Acq On : 5 Oct 2018 14:03
 Operator : MD\SY
 Sample : J5239-01 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(AUG)

Quant Time: Oct 05 15:30:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N100518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 05 11:11:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	95723	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	530024	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	440559	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	195808	29.47	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	98.23%
60) 4-Bromofluorobenzene	12.40	95	150611	22.40	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	74.67%
63) Toluene-d8	10.09	98	624517	31.93	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	106.43%

Target Compounds

					Ovalue
15) Acetone	3.82	58	44868	130.407	ug/l 80
25) Chloroform	7.37	83	27127	2.626	ug/l # 81
41) Bromodichloromethane	9.40	83	2780	0.345	ug/l # 85
62) Toluene	10.16	91	306504	13.734	ug/l 98
82) p-Isopropyltoluene	13.29	119	11183	0.467	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN100518\
Data File : VN051684.D
Acq On : 5 Oct 2018 14:03
Operator : MD\SY
Sample : J5239-01 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(AUG)

Quant Time: Oct 05 15:30:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N100518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 05 11:11:26 2018
Response via : Initial Calibration

